

Certificate



Certificate no.

TA 50422408 01

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

030-0000334704

Test report no.:

ZTW1-SCL 50144994 003

Client Reference:

EZ-18102018-Co.MW

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-11:2015
EN 60601-1-11:2015

Certified Product:

(Switching Power Adapter)

License Fee - Units

Type Designation	: GSM60Ebzzzzz (MW)	5
b stands for	: 05, 07, 09, 12, 15, 18, 24, 28, 48 or 12A	1
z stands for	: 0 to 9, A to Z, hyphen or blank	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1.5-0.8A	
Output Voltage	: see Appendix	
Output Current	: see Appendix	
max. Ambient Temperature	: 40°C	
Protection Class	: II	

Remark: 1. The usability requirements of the standard were not addressed.
2. This certificate is valid for maximum 5 years.

7

Licensed Test mark:



IEC 60601-1
EN 60601-1
IEC 60601-1-11
EN 60601-1-11

www.tuv.com
ID 1419063683

TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

13/11/2018

Certificate



Certificate no.

TA 50422408 02

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

030-0000334704

Test report no.:

ZTW1-SCL 50144994 003

Client Reference: EZ-18102018-Co.MW

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-11:2015
EN 60601-1-11:2015

Certified Product: (Switching Power Adapter)

License Fee - Units

as page 01

Addition

Type Designation	: GEM60Ibzzzzz (MW)	1
b stands for	: 05, 07, 09, 12, 15, 18, 24, 28, 48 or 12A	1
z stands for	: 0 to 9, A to Z, hyphen or blank	1

This certificate is valid for maximum 5 years.

3

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

13/11/2018

Certificate



Certificate no.

TA 50422408 03

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu District,
New Taipei City 24891
Taiwan

Manufacturing Plant:

040-0002002959

Test report no.: ZTW1-WY 50144994 008

Client Reference: P01121312

Tested to:

IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2
IEC 60601-1-6:2010+A1+A2
EN 60601-1-6:2010+A1+A2
IEC 62366-1:2015+A1
EN 62366-1:2015+A1
IEC 60601-1-11:2015+A1
EN 60601-1-11:2015+A1

Certified Product: (Switching Power Adapter)

License Fee - Units

as page 01

Change

License Holder : see above
Test Requirement : see above

This certificate is valid for maximum 5 years.

Licensed Test mark:



IEC 60601-1
EN 60601-1
IEC 60601-1-11
EN 60601-1-11

www.tuv.com
ID 1419063683

TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. (FH) A. Klinker



Date of Issue
(day/mo/yr)

03/11/2023

Appendix to TÜV approved Certificate No.: TA 50422408

Certified Product : **Switching Power Adapter**
Report Number : **50144994 003**
Type Designation : **GSM60Ebzzzzz, GSM60Ibzzzzz**
(b can be 05, 07, 09, 12, 15, 18, 24, 28, 48 or 12A;
z can be 0 to 9, A to Z, hyphen or blank)

Model name	Output Voltage (Vdc)	Output Current (A)	Max. output power (W)
GSM60E05zzzzz, GEM60I05zzzzz	5-6	6.00-5.00	30
GSM60E07zzzzz, GEM60I07zzzzz	6-8	6.00-5.63	45.04
GSM60E09zzzzz, GEM60I09zzzzz	8-11	5.63-5.00	55
GSM60E12Azzzzz, GEM60I12Azzzzz	11-13	5.0-4.62	60
GSM60E12zzzzz, GEM60I12zzzzz	12	4.5	54
GSM60E15zzzzz, GEM60I15zzzzz	13-16	4.62-3.75	60
GSM60E18zzzzz, GEM60I18zzzzz	16-21	3.75-2.86	60.06
GSM60E24zzzzz, GEM60I24zzzzz	21-27	2.86-2.23	60.21
GSM60E28zzzzz, GEM60I28zzzzz	27-33	2.23-1.82	60.06
GSM60x48zzzzz, GEM60I48zzzzz	33-58	1.82-1.04	60.32

Date: 13th November, 2018



Certification Body

Dipl.-Ing. (FH) A. Klinker