

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

## Product

Switching Power Supply

## Name and address of the applicant

MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan

## Name and address of the manufacturer

MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan

## Name and address of the factory

MEAN WELL Enterprises Co., Ltd.  
No.28, Wuquan 3rd Rd, Wugu District, 24891 New Taipei City, Taiwan*When more than one factory*☒ Additional information on page 3

## Ratings and principal characteristics

## Product data

Ratings: Input: 100 -240 Vac, 1,8-0,7 A, 50/60 Hz or  
277 Vac, 0,63 A, 50/60 Hz.

Output: See General product information of report for details.

## Product data – type NSP-150-12zzzz

Rated output voltage: 12,0 Vdc; Rated output current: 12,5 A

## Product data – type NSP-150-15zzzz

Rated output voltage: 15,0 Vdc; Rated output current: 10,0 A

## Product data – type NSP-150-24zzzz

Rated output voltage: 24,0 Vdc; Rated output current: 6,3 A

## Product data – type NSP-150-27zzzz

Rated output voltage: 27,0 Vdc; Rated output current: 5,6 A

## Product data – type NSP-150-36zzzz

Rated output voltage: 36,0 Vdc; Rated output current: 4,2 A

## Product data – type NSP-150-48zzzz

Rated output voltage: 48,0 Vdc; Rated output current: 3,15 A

## Product data – type NSP-150-5zzzz

Rated output voltage: 5,0 Vdc; Rated output current: 30,0 A

## Product data – type NSP-150-60zzzz

Rated output voltage: 60,0 Vdc; Rated output current: 2,55 A

## Product data – type NSP-150-7.5zzzz

Rated output voltage: 7,5 Vdc; Rated output current: 20,0 A

## Trademark / Brand (if any)



MEAN WELL

## Customer's Testing Facility (CTF) Stage used

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V.  
Meander 1051  
6825 MJ Arnhem  
Netherlands

Signature: Rosa Zhou



Date: 2025-11-28

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

Model / Type Ref.

NSP-150-12zzzz, NSP-150-15zzzz, NSP-150-24zzzz, NSP-150-27zzzz,  
NSP-150-36zzzz, NSP-150-48zzzz, NSP-150-5zzzz, NSP-150-60zzzz and  
NSP-150-7.5zzzz

Additional information

☒ Additional information on page 3A sample of the product was tested and  
found to be in conformity withIEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and  
IEC 60601-1:2005/AMD2:2020

## National differences:

CA, CH, CN, EU Group Differences, EU Special National Conditions, GB,  
JP, KR, USAs shown in the Test Report Ref. No.  
which forms part of this Certificate

4940540.50

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V.  
Meander 1051  
6825 MJ Arnhem  
Netherlands

Date: 2025-11-28

Signature: Rosa Zhou

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME****Additional factory**

MEAN WELL (GUANGZHOU) Electronics Co., Ltd.  
No.11 Jingou South Road, Huadu District, 510890 Guangzhou Guangdong, China

YONGDEN TECHNOLOGY CORPORATION  
345 MacArthur Highway, Tabang, 3015 Guiginto, Bulacan, Philippines

SuZhou MEAN WELL Technology Co., Ltd.  
No.269, Changping Road, Huangdai Town, Xiangcheng District, 215152 Suzhou Jiangsu, China

MEAN WELL INDIA ELECTRONICS PRIVATE LIMITED  
9C, Peenya industrial area, Chokkasandra, 2ND PHASE, PEENYA, 560058 Bengaluru (BANGALORE) Urban,  
Karnataka, India

**Additional information**

Model/Type reference:  
NSP-150-xzzzz  
(x=5, 7.5, 12, 15, 24, 27, 36, 48, 60; z=may be blank. -, 0~9, A~Z,  
a~z for market purpose)

Only the requirements of IEC 60601-1, as documented in the accompanying test report, have been addressed.

This CB Test Certificate is issued by the National Certification Body

DEKRA Certification B.V.  
Meander 1051  
6825 MJ Arnhem  
Netherlands



Date: 2025-11-28

Signature: Rosa Zhou